



PERFORMANCE VERIFICATION STATEMENT For Systea WIZ Probe Phosphate Analyzer

| | |
|------------------------------|---|
| TECHNOLOGY TYPE: | Nutrient Sensors |
| APPLICATION: | In situ estimates of PO ₄ for coastal moored deployments |
| PARAMETERS EVALUATED: | Accuracy, precision, range response and reliability |
| TYPE OF EVALUATION: | Laboratory and Field Performance Verification |
| DATE OF EVALUATION: | Testing conducted from January 2015 to November 2016 |
| EVALUATION PERSONNEL: | T. Johengen, H. Purcell, M. Tamburri, D. Loewensteiner, G.J. Smith, D. Schar, Margaret McManus, Gordon Walker, and Beth Stauffer. |

NOTICE:

ACT verifications are based on an evaluation of technology performance under specific, agreed-upon protocols, criteria, and quality assurance procedures. ACT and its Partner Institutions do not certify that a technology will always operate as verified and make no expressed or implied guarantee as to the performance of the technology or that a technology will always, or under circumstances other than those used in testing, operate at the levels verified. ACT does not seek to determine regulatory compliance; does not rank technologies nor compare their performance; does not label or list technologies as acceptable or unacceptable; and does not seek to determine “best available technology” in any form. The end user is solely responsible for complying with any and all applicable federal, state, and local requirements.

This document has been peer reviewed by ACT Partner Institutions and a technology-specific advisory committee and was recommended for public release. Mention of trade names or commercial products does not constitute endorsement or recommendation by ACT for use.

Questions and comments should be directed to: Dr. Tom Johengen
ACT Chief Scientist
CILER- University of Michigan
4840 S. State Street
Ann Arbor, MI 48108 USA
Email: Johengen@umich.edu

TABLE OF CONTENTS

| | |
|--|-----------|
| EXECUTIVE SUMMARY | 3 |
| BACKGROUND AND OBJECTIVES | 5 |
| INSTRUMENT TECHNOLOGY TESTED | 5 |
| PERFORMANCE EVALUATION TEST PLAN..... | 6 |
| LABORATORY TESTS..... | 6 |
| FIELD TESTS | 8 |
| REFERENCE SAMPLE ANALYSIS..... | 10 |
| RESULTS OF LABORATORY TEST..... | 12 |
| RESULTS OF FIELD TESTS..... | 22 |
| DEPLOYMENT AT MAUMEE RIVER BOWLING GREEN, OHIO | 23 |
| DEPLOYMENT AT CHESAPEAKE BIOLOGICAL LABORATORY | 28 |
| DEPLOYMENT OFF COCONUT ISLAND IN KANEOHE BAY, HAWAII | 33 |
| QUALITY ASSURANCE/QUALITY CONTROL | 39 |
| ACKNOWLEDGEMENTS | 45 |

EXECUTIVE SUMMARY

The Alliance for Coastal Technology (ACT) conducted a sensor verification study of in situ nutrient analyzers during 2016 to characterize performance measures of accuracy, precision and reliability. The verification including a week of laboratory testing along with three moored field deployments in freshwater, estuarine, and oceanic coastal environments. Laboratory tests of accuracy, precision, and range were conducted at the University of Maryland's Chesapeake Biological Laboratory (CBL) in Solomons, MD. A series of five tests were conducted to evaluate performance under controlled challenge conditions including: concentration range, temperature, salinity, turbidity, and dissolved organic carbon. All laboratory tests were conducted in 250 L polypropylene tanks using RO water as the initial matrix, within a temperature controlled room. Instruments sampled from a common, well-mixed, test tank maintained at a documented level of known challenge condition. Instruments were set-up by the manufacturer daily prior to the start of each individual laboratory tests, exposed to each test condition for a period of three hours, and programmed to sample at a minimum frequency of 30 minutes. Reference samples were collected every 30 minutes for five timepoints during corresponding instrument sampling times for each test.

For the laboratory concentration range challenge the absolute difference between the Systea-PO4 and reference measurement across all timepoints for trials C0 – C5 ranged from -0.0067 to 0.0277 mgP/L, with a mean of 0.0041 ± 0.0091 mgP/L. There was a small but significant increase in the measurement difference with increasing concentration as determined by linear regression ($p=0.02$; $r^2=0.20$). However, the change in accuracy mostly occurred at the highest two test concentrations (0.406 and 1.87 mgP/L) with absolute measurement errors of 0.022 and 0.171 mgP/L, respectively. An assessment of precision was performed by computing the standard deviations and coefficients of variation of the five replicate measurements for C1 – C5 concentration trials. The standard deviation of the mean ranged from 0.0005 to 0.0040 mgP/L across the five trials, and the coefficient of variation ranged from 0.93 to 4.75 %. For the laboratory temperature challenge with testing at 5 °C, the absolute difference between instrument and reference measurement across all timepoints for trials C2 – C4 ranged from -0.0026 to 0.0170 mgP/L, with a mean of 0.010 ± 0.005 mgP/L. Measurement differences were significantly higher at 5 °C versus 20 °C for each of three concentration test with the increased offset being 0.007, 0.005 and 0.017 mgP/L greater, respectively, for C2, C3, and C4. Similar to test results at 20 °C, there was a much larger offset at the C4 level compared to the lower two concentrations for the 5 °C test. For the laboratory salinity challenge performed at the C3 concentration level, the absolute difference between instrument and reference measurement across all timepoints for the three added salinity levels ranged from 0.0030 to 0.0333 mgP/L, with a mean of 0.0205 ± 0.0078 mgP/L. There was a statistically significant response to increased salinity with measurement offsets increasing (over-predicted) as salinity increased. A linear regression of the measurement differences versus salinity was significant ($p<0.001$; $r^2=0.86$) with a slope of 0.0009 and intercept of -0.0010. The average offset at salinity 30 was around 0.028 mgP/L higher than for the zero salinity trial which corresponded to a relative error of approximately 90%. For the laboratory turbidity challenge, performed at the C3 concentration level, the absolute difference between instrument and reference measurement across all timepoints for the two added turbidity levels ranged from -0.0208 to 0.0046 mgP/L, with a mean of -0.0079 ± 0.0091 mgP/L. A linear regression of the measurement differences versus turbidity was significant ($p<0.01$; $r^2=0.59$), with a slope of -0.0008 and intercept of 0.004, however the trend line was clearly forced by the large decrease at 100 NTU and test results should not be interpreted to suggest a strong predictable relationship. For the laboratory DOC challenge, performed at the C3 concentration level, the

absolute difference between instrument and reference measurement across all timepoints for the two added DOC levels ranged from -0.0018 to 0.0100 mgP/L, with a mean of 0.0018 ± 0.0034 mgP/L. A linear regression of the measurement differences versus DOC concentration was not significant ($p=0.71$; $r^2=0.01$).

A 32 day deployment occurred from May 26 through June 27 in the Maumee River, at the facilities of the Bowling Green, Ohio Water Treatment Plant. The Systea-PO4 operated successfully during the entire 32 day deployment sampling at 15 minute intervals. The Systea-PO4 generated 3031 observations out of a possible 3052 for a data completion result of 99.3%. During the deployment 21 data points were not reported by the instrument. The average and standard deviation of the measurement difference between instrument and reference PO₄ measurements for each matched pair ($n=50$ of a possible 51 observations) over the total deployment was 0.004 ± 0.024 mgP/L with a total range of -0.086 to 0.035 mgP/L. There was no significant trend in measurement difference over time as estimated by linear regression ($p=0.86$; $r^2=0.001$). A linear regression of instrument versus reference measurement was significant ($p<0.01$; $r^2=0.25$) but with a slope of only 0.28 and intercept of 0.019. Measurement accuracy clearly declined when concentrations exceeded 0.05 mgP/L.

An 84 day moored field test was conducted in Chesapeake Bay from July 18 to October 10, 2016. The Systea-PO4 operated continuously for the period of its deployment sampling at 30 minute intervals but was retrieved 12 days prior to the scheduled end date of the deployment to send to the next field test in HI. While the unit was deployed it reported 3086 of a possible 3402 accepted values for a data completion result of 90.7%. During its operation, 64 values were flagged by the instrument with no data, and 252 were omitted as outliers when reported values exceeded 10 times above observed levels or were less than -0.004. The average and standard deviation of the measurement difference between instrument and reference PO₄ measurements for each matched pair ($n=78$ of a possible 103 observations) over the total deployment was 0.019 ± 0.014 mgP/L, with the total range of differences between -0.019 to 0.049 mgP/L. There was no significant trend in measurement difference over time as estimated by linear regression ($p=0.58$; $r^2=0.004$) over the deployment period. A linear regression of the instrument versus reference measurements was not significant ($p=0.58$; $r^2=0.004$) and the instrument generally over-predicted concentrations.

A one month long moored field test was conducted in Kaneohe Bay from October 3, 2016 to November 2, 2016. The Systea-PO4 was deployed on day four of the field test and operated successfully for the remaining 26 days of the deployment. During the deployment the Systea-PO4 returned 2477 acceptable instrument measurements of a possible 2496 measurements for a data completion result of 99% (13 data points were not reported and 6 observations were omitted as outliers having values >10x maximum reference). The average and standard deviation of the differences between instrument and reference readings over the entire deployment ($n=63$ out of a possible 63) were -0.0004 ± 0.002 mgP/L, with a total range of -0.0058 to 0.0044 mgP/L. There was a small but statistically significant trend in the measurement difference over time ($p=0.0003$; $r^2=0.247$) during the deployment, with a slope of 0.0001 mgP/L/d. A linear regression of the instrument versus reference data was highly significant but with a low regression coefficient ($p=0.001$; $r^2=0.209$). The regression had a slope of 0.999 and y intercept of -0.0004. Under and over predictions were fairly evenly distributed.

BACKGROUND AND OBJECTIVES

The Alliance for Coastal Technologies (ACT) is a NOAA and EPA-funded partnership of research institutions, state and regional resource managers, and private sector companies that are interested in developing, improving, and applying sensor technologies for studying and monitoring coastal environments. ACT was established on the premise that instrument validation of existing and emerging technologies is essential to support both coastal science and resource management. The overall goals of ACT's verification program are to provide industry with an opportunity to have a third-party test their instruments in both controlled laboratory settings and in diverse field applications within a range of coastal environments, and to provide users of this technology with an independent and credible assessment of instrument performance.

ACT partnered with the multi-agency Challenging Nutrients Coalition on the Nutrient Sensor Challenge to help address the environmental and ecological problems associated with nutrient pollution. A critical step in this process is facilitating the development and adoption of the next-generation of *in-situ* nutrient sensors and analyzers. To that end, the ACT Technology Verification model was applied to the Nutrient Sensor Challenge to test instrument performance in laboratory and field tests against reference water samples analyzed using EPA-approved standard methods.

The report within contains the test results for the Systea WIZ Phosphate Analyzer during the ACT Performance Verification. A synthesis of the testing protocols and reference sample analysis are provided below. A complete copy of the verification protocols is available on the ACT website at the following link: http://www.act-us.info/nutrients-challenge/Download/Nutrient_Challenge_Test%20Protocols_PV16_01.pdf

INSTRUMENT TECHNOLOGY TESTED

The WIZ probe is the latest development of Systea, a state of the art portable "in-situ" probe that measures phosphate continuously in surface waters or marine environments. The innovative design allows easy handling and field deployment by the user. The WIZ probe allows, in the standard configuration, the detection at trace levels of four nutrient parameters (orthophosphate, ammonia, nitrite and nitrate). The WIZ probe autonomously manages the well tested spectrophotometric wet chemistries as well as an advanced fluorometric method for ammonia measurement.

For this evaluation a mono-parametric probe that measures orthophosphate was tested. The Systea WIZ Phosphate Analyzer measures orthophosphate using acid molybdate solution and ascorbic acid. The blue color of the phosphor-molybdenum complex formed is measured spectrophotometrically at 880 nm. With this method the WIZ is able to cover the range from 2 ppb to 1500 ppb without dilution.

The probe uses the micro Loop Flow Analysis (μ LFA) that is an analytical technology for autonomous management of a microfluidic system to handle complex analytical methods using a batch principle. In a 1.5 ml volume hydraulic loop, the water sample is collected and the required reagents are sequentially injected and mixed to perform the specific conditioning procedure needed for an analytical reaction. As soon as the measurement is performed, the hydraulic circuit is washed with DI water. The process can be repeated again with the same method or with a different analytical procedure. The small reactor enables an extremely low consumption of reagents and calibrants permitting the design of a "plug-in" compact reagent container to allow an immediate

field reagent and calibration solutions changeover, ensuring real field portability; it can contain up to 1000 ml of solutions in several flexible bags.

Results are directly provided in concentration units; all measured values are stored with date, time and sample optical density (O.D.). The same data are remotely available through a serial communication port, which allows the complete probe configuration and remote control using the external Windows® based Wiz Control Panel software.

PERFORMANCE EVALUTION TEST PLAN

These Test Protocols are based on consensus recommendations of the ACT Technical Advisory Committee, ACT staff, and participating Manufacturers. In summary, the test:

- utilized standard, approved laboratory analytical methods to provide best possible measure of the ‘true’ nutrient concentration from reference samples, which served as performance standards against which instrument estimations were compared internally by the individual developer;
- conducted all reference sample analysis at the state certified Nutrient Analytical Services Laboratory (NASL) of the Chesapeake Biological Laboratory (CBL), Solomons, MD to determine true nutrient concentrations using USEPA approved methodologies (see details below);
- included a laboratory evaluation of instrument performance;
- included three moored/dock-based field trials under a wide range of environmental conditions including freshwater, estuarine and marine ecosystems with varying nutrient concentrations and water quality characteristics (e.g. turbidity).

All ACT personnel involved in the Nutrient Sensor Verification were trained on standardized water sample collection, storage and shipping methods. ACT staff was available to assist in the physical deployment and recovery of all submitted test instruments and were responsible for the data management of test instrument results. Challenge participants were responsible for initial set-up and calibration of their instrument. If requested, ACT provided the chemicals and nutrient standards needed for instrument set-up and calibration. All laboratory nutrient analyses of the independent reference samples were conducted at the CBL NASL using standardized automated wet chemistry. All numerical data were recorded to three significant decimals where appropriate and nutrient concentrations reported in elemental mass units as mgN/L or mgP/L for nitrate+nitrite (NO_{23}), nitrate (NO_3^-) or phosphate (PO_4^{3-}), respectively.

Laboratory Tests

Laboratory tests of accuracy, precision, and range were conducted at the University of Maryland’s Chesapeake Biological Laboratory (CBL) in Solomons, MD. A series of five tests were conducted to evaluate performance under controlled challenge conditions including: concentration range, temperature, salinity, turbidity, and dissolved organic carbon (details below). All Laboratory tests were conducted in polypropylene tank using RO water as the initial matrix, within a temperature controlled room. All instruments sampled from a common, well-mixed, test tank of approximately 250L volume, maintained at a documented level of known challenge condition. Instruments were set-up by the manufacturer daily prior to the start of each individual laboratory tests. Instruments were exposed to each test condition for a period of three hours and programmed to sample at a minimum frequency of 30 minutes. Reference samples were collected

every 30 minutes for five timepoints during instrument sampling times for each test. Laboratory tests included the following ‘controlled’ challenge conditions:

Test 1: Accuracy and Precision over a broad concentration range

- Tested response across a broad range of concentrations representative of natural waters.
 - Concentration levels for NO₃ (mgN/L): 0.005, 0.1, 1.0, 5, 10, and 50
 - Concentration levels for PO₄ (mgP/L): 0.002, 0.01, 0.05, 0.1, 0.5, and 2.0
- The range test was split into two separate tests with concentrations for levels 1-4 conducted on day 1 and the last two concentrations tested on day 6 due to time constraints. Note that the starting level on day 6 was mistakenly set to 5 mgN/L and the 10 mgN/L level was not actually tested.
- Three hour sampling windows were provided at each of the six concentrations during which instruments measured concentrations at a minimum frequency of every 30 minutes.
- Discrete reference samples were collected every 30 minutes, corresponding to instrument sampling times, to generate five comparative measurements to assess accuracy and precision against reference values.
- RO water was used as the test matrix to which known amounts of nutrient salts (KNO₃ and K₂HPO₄) were added. Analysis of ambient blanks indicated a small amount of inorganic nutrients in the RO water.
- Tests were conducted at 20 °C in a temperature controlled room with samples drawn from a common well-mixed 250L test tank.

Test 2: Temperature Response

- Instrument response was tested for three concentrations, corresponding to levels C2, C3, and C4 from the range test, at temperatures of 5 °C versus the temperature of 20 °C on the first day.
- Temperature was regulated and maintained within a temperature controlled room and independently verified in the test tank with an YSI EXO2 reading at 15 min intervals.
- Instruments were equilibrated to the new 5 °C test temperature overnight.
- Instruments were exposed for three hours at each of the 3 concentrations with reference samples collected every 30 minutes following an initial 30 minute equilibration period to each condition.

Test 3: Salinity Response

- Accuracy and precision was tested over three additional salinities (10-20-30) at the C3 concentration level of the range test at 20°C.
- Salinity levels were developed using Instant Ocean additions to the RO water matrix, which could have contributed trace amounts of nutrients, but would have measured in the final reference samples.
- Instruments were exposed for three hours at each of the 3 concentrations with reference samples collected every 30 minutes following an initial 30 minute equilibration period to each condition.

Test 4: Turbidity Response

- Accuracy and precision were tested over two elevated turbidity levels (approximately 10 and 100 NTU) at the C3 concentration level of the range test at 20 °C.
- Test tanks were continuously mixed with submersed pumps but there was some settling of the material as noted by continuous monitoring with the EXO2 sonde and analysis of discrete turbidity samples on the Hach 2100.
- Turbidity concentrations were established using Elliot Silt Loam reference material (cat # 1B102M) available from the International Humic Substances Society (<http://www.humic-substances.org>) added into RO water matrix.
- Instruments were exposed for three hours at each of the 3 concentrations with reference samples collected every 30 minutes following an initial 30 minute equilibration period to each condition.

Test 5: DOC Response

- Accuracy and precision were tested against two DOC levels (1 and 10 mg/L) at the C3 concentration level of the range test at 20 °C.
- DOC concentrations were established using the Upper Mississippi River Natural Organic Matter standard (cat# 1R110N) available from the International Humic Substances Society (<http://www.humic-substances.org>) added to RO water matrix.
- Instruments were exposed for three hours at each of the 3 concentrations with reference samples collected every 30 minutes following an initial 30 minute equilibration period to each condition.

Field Tests

In situ field performance evaluations of the test instruments were conducted under extended mooring deployments at three ACT Partner Institution sites covering freshwater, estuarine, and marine conditions. Site specific details for each test site were as follows:

Freshwater Deployment: The freshwater deployment occurred on the Maumee River in Waterville, OH for one month duration and provided a high nutrient, high turbidity test environment. The ACT Partner at the University of Michigan established a flow-through system on the Maumee River near Waterville Ohio (83.74 °N; 41.48 °W), located within the pump house of the City of Bowling Green Municipal Water Treatment Plant. Instruments were deployed in a 180 gallon flow-through tank with a water depth of approximately 0.8m and exchange time of approximately 10 minutes. The Maumee River main stem flows 137 km before flowing into the Maumee Bay of Lake Erie at the city of Toledo, Ohio. The Maumee watershed is the largest watershed of any Great Lakes river with 8,316 square miles. The majority of the watershed is cultivated crop land, mostly corn and soybeans, though concentrated areas of pasture are located in the northwestern and southeastern areas of the watershed.

Estuarine deployment : The estuarine deployment occurred at the research pier of the Chesapeake Biological Laboratory in Solomons, MD for three month duration and provided for variable salinity and nutrient levels within a highly productive and biofouling environment. The ACT Partner at Chesapeake Biological Laboratory (CBL), University of Maryland Center for Environmental Science, has established a Technology Verification Field Test Site on a fixed pier (38.32 °N;76.45 °W), with an average depth of 2.1 m at the mouth of the Patuxent River, a tributary of the Chesapeake Bay. The deployment frame was arranged so that all of the sample

inlets for the instruments remain at a fixed depth of 1 m below the water surface using a floating dock. The Chesapeake is a nutrient rich estuary with a watershed that encompasses portions of six states and the District of Columbia. Water temperatures at the test site ranged from 20 to 31°C and salinity ranged from 12.7 to 16.9 psu during the Verification.

Marine deployment: The marine deployment occurred in Kaneohe Bay at the Hawaii Institute of Marine Biology field lab for one month duration and provided a full salinity, low nutrient test condition. The ACT Partner at the Hawaii Institute of Marine Biology (HIMB) is part of the University of Hawaii with a field site established on the Kaneohe Bay Barrier Reef flat (21.43 °N; 157.79 °W) in waters ~16 m deep. The deployment frame was arranged so that all of the sample inlets for the instruments remain at a fixed depth of 1 m below the water surface using a floating dock. Kaneohe Bay sits on the northeast, or windward, side of Oahu. Water temperatures at this site varied between 24.5 and 27.9°C and salinities were between 27.3 and 34.8 psu during the Verification.

Instrument Setup - Prior to deployment, all instruments were set up and calibrated as required at the field sites by a manufacturer representative, with assistance provided by ACT staff as necessary. The manufacturer supplied or specified to ACT all specific materials and hardware (chemicals, power cords, cables, weights, etc.) needed to deploy the test instrument according to requirements defined for each field site. ACT staff worked with the manufacturer to design an appropriate sensor deployment configuration at each site and arranged instruments in a manner so that a single representative field sample could be collected without the potential of interference between instruments. No servicing of the instruments was to occur during the test deployment period unless observed physical damage had occurred from natural events and a repair or replacement was deemed necessary. Instruments were set up as self-recording, either internally or to an external data logger, and programmed to record data based on a time interval that allowed instruments to function for the specified number of days for the respective deployment. Specific sampling intervals varied among test instruments, but with a stated goal of 15 minute sampling intervals if possible and two-hour intervals at maximum. A sampling schedule was established so that all instruments being tested at the same time had a common sampling time point at a minimum frequency of 2 hours. Internal clocks were set to local time and synchronized against the time standard provided by www.time.gov.

Reference Water Sampling Schedule – The reference sampling schedule generated between 50 - 100 comparative reference samples and was structured to examine changes in nutrient concentrations over daily to monthly time scales. Specifically, once each week ACT staff conducted an intensive sampling event that consisted of four consecutive samples spaced at two-hour intervals. For the remaining four days of the week, ACT staff sampled once or twice per day, spaced out to cover early morning and late-afternoon timepoints or anticipated flow or tidal events. The initial intensive sampling event occurred within the first two days of the deployment after all instruments had been deployed, and the final intensive sampling event occurred during the last two days of the deployment.

Reference Water Sample Collection - A standard 2L Van Dorn bottle was used at the CBL and HI field sites to collect reference water samples for laboratory nutrient concentration analysis. For the riverine test site a 1L acid-cleaned, polypropylene bottle was filled directly from the flow-through tank. For the tank sampling, the sampling bottle was rinsed three times before filling. For the mooring sites, the Van Dorn bottle was lowered to the same depth and as close as physically possible to the sampling inlets of all instruments and less than 1 m from any individual sampling

inlet and soaked at sampling depth for 1 minute prior to sampling. The water sample was then transferred to an acid washed 1L polypropylene bottle after three initial rinses of the field sample. All environmental reference samples were processed within 10 minutes of collection while wearing clean laboratory gloves to minimize potential sources of contamination. The sample was filtered through a 47mm Whatman GFF filter into an acid cleaned vacuum flask. The first 50 ml of filtrate were discarded as a rinse. The remaining filtrate was distributed into 8 individual acid-cleaned, 30 ml polypropylene bottles to provide three analytical replicates each for NO₃ and PO₄ plus two replicates to hold as back-ups. All final sample bottles were rinsed once before filling and filled no more than ¾ full to allow adequate headspace for freezing. The final reference samples were immediately frozen and remained so until shipment to CBL-NASL for analysis.

Sample Handling and Chain of Custody - All collected reference samples at each test site were dated and coded according to site and sample sequence. Each sample container was labeled with a number for identification. The reference sample number was used in all laboratory records and Chain-of-Custody (COC) forms to identify the sample. Samples were shipped on dry ice to CBL-NASL for nutrient analysis within approximately two weeks of collection. Shipping containers were sent for next morning delivery, or the soonest possible delivery time possible from a given shipping location. All samples, including the condition shipped and received, were recorded onto Chain of Custody (COC) forms and a copy sent with the samples. The COC specified time, date, sample location, unique sample number, requested analyses, sampler name, and required turnaround time, time and date of transaction between field and laboratory staff, and name of receiving party at the laboratory. NASL confirmed receipt and condition of samples within 24 hours of their arrival by signing and faxing a copy of the form to the test site.

Reference Sample Analysis

Phosphate concentrations for all reference and quality control samples were determined by the NASL at CBL following their Standard Operating Procedures Manual (CEES, UMD, Publication Series No. SS-80-04-CBL). The methodology is based on U.S. EPA Method 365.1, in Methods for chemical analysis of water and wastes (United States Environmental Protection Agency, Office of Research and Development. Cincinnati, Ohio. Report No. EPA-600-4-79-020 March 1979). In brief, ammonium molybdate and antimony potassium tartrate react in an acidic medium with dilute solutions of phosphate to form an antimony-phospho-molybdate complex. The complex is reduced to an intensely blue-colored complex by ascorbic acid. The color produced is proportional to the phosphate concentration present in the sample.

Nitrate and nitrite concentrations for all reference and quality control samples were determined by the NASL at CBL following their Standard Operating Procedures Manual (CEES, UMD, Publication Series No. SS-80-04-CBL). The methodology is based on U.S. EPA Method 353.2, in Methods for chemical analysis of water and wastes (United States Environmental Protection Agency, Office of Research and Development. Cincinnati, Ohio. Report No. EPA-600-4-79-020 March 1979). In brief, nitrate is reduced to nitrite using the cadmium reduction method. The nitrite is then determined by diazotizing with sulfanilamide and coupling with N-1-naphthylethylenediamine di hydrochloride to form a color azo dye. The absorbance measured at 540 nm is linearly proportional to the concentration of nitrate + nitrite in the sample. Nitrate concentrations are obtained by subtracting nitrite values, which have been separately determined without the cadmium reduction procedure.

All laboratory nutrient analyses were conducted on an Aquakem 250 auto-analyzer. For

phosphates, a statistically-determined method of detection limit for this instrument of 0.0007 mgP/L was established by prior laboratory studies for a wide range of salinities. An expected working concentration range for this Verification and SOP was between 0.002 and 1.48 mgP/L. The detection limits for nitrate and nitrite were similarly established at 0.0007 mgN/L and 0.0006 mgN/L respectively. The typical working concentration range for the nitrate method and SOP is between 0.0049 – 5.6 mgN /L. The typical working concentration range for the nitrite method and SOP is between 0.0042 – 0.28 mgN /L. The system contains an auto-dilutor to bring any higher concentrations down to the established linear calibration range. A sample reagent blank is analyzed in conjunction with every sample as part of the routine operation of the Aqaukem 250. Approximately 40 samples per hour can be analyzed. All internal standards were verified and calibrated using certified external nutrient standards (such as Spex Certi-Prep or NIST). In addition, Field Trip Blanks and Field Sample Spike Additions (defined below) were conducted once per week by ACT as part of established quality assurance/quality control (QA/QC) protocols.

RESULTS OF LABORATORY TEST

Accuracy

Systea-PO4 measurements and corresponding reference measurements for the lab concentration range challenge are shown in figure 1. Results for the highest concentration are excluded from any numerical or statistical comparisons because of its extreme range, but were included in the test to help identify maximum detection potential. The absolute difference between instrument and reference measurement across all timepoints for trials C0 – C5 ranged from -0.0067 to 0.0277 mgP/L, with a mean of 0.0041 ± 0.0091 mgP/L (Table 1). A plot of the absolute difference between Systea-PO4 and reference measurement is shown in the bottom panel of figure 1. There was a small but significant increase in the measurement difference with increasing concentration as determined by linear regression ($p=0.02$; $r^2=0.20$). However, the change in accuracy mostly occurred at the highest two test concentrations (0.406 and 1.87 mgP/L) with measurement errors of 0.022 and 0.171 mgP/L, respectively.

Table 1. Accuracy results for laboratory testing of the Systea-PO4 analyzer assessed by absolute difference (mgP/L) and percent error between instrument and reference measurements for the concentration range test.

| Trial | Reference | Systea-PO4 | Absolute Diff | % Error |
|-------|-----------|------------|----------------|-------------|
| C0 | 0.0059 | 0.0058 | -0.0001 | 1.5 |
| C1 | 0.0105 | 0.0141 | 0.0036 | 33.8 |
| C2 | 0.0189 | 0.0212 | 0.0023 | 12.2 |
| C3 | 0.0621 | 0.0619 | -0.0002 | 0.3 |
| C4 | 0.1159 | 0.1116 | -0.0043 | 3.7 |
| C5 | 0.4059 | 0.4274 | 0.0215 | 5.3 |

Precision

An assessment of precision was performed by computing the standard deviations and coefficients of variation of the five replicate measurements for C1 – C5 concentration trials. The standard deviation of the mean ranged from 0.0005 to 0.0040 mgP/L across the five trials, and the coefficient of variation ranged from 0.93 to 4.75 % (Table 2).

Table 2. Precision assessment of the Systea-PO4 analyzer during the laboratory concentration range testing. Variance is reported as the standard deviation and coefficient of variation of five replicate measurements collected at 30 minute intervals in a well-mixed tank maintained at known uniform conditions.

| Trial | Mean PO ₄ (mgP/L) | | Standard Deviation | | Coefficient of Variation | |
|-------|------------------------------|------------------------|--------------------|------------------------|--------------------------|------------------------|
| | Reference | Systea-PO ₄ | Reference | Systea-PO ₄ | Reference | Systea-PO ₄ |
| C1 | 0.0105 | 0.0141 | 0.0004 | 0.0007 | 4.21 | 4.75 |
| C2 | 0.0189 | 0.0212 | 0.0001 | 0.0005 | 0.66 | 2.34 |
| C3 | 0.0621 | 0.0619 | 0.0005 | 0.0013 | 0.75 | 2.04 |
| C4 | 0.1159 | 0.1116 | 0.0003 | 0.0021 | 0.23 | 1.86 |
| C5 | 0.4059 | 0.4274 | 0.0023 | 0.0040 | 0.56 | 0.93 |

Lab Concentration Range Challenge




Figure 1. *Top Panel:* Plot of instrument (blue dots) and reference (red dots) measurements of PO₄ in the laboratory concentration range challenge covering ambient plus 6 concentration ranges. Five replicate measurements were made at each concentration level along with three measurements at ambient level. *Bottom Panel:* Plot of the absolute difference in mgP/L between Systea-PO4 and reference measurement.

Time series results of ambient water quality conditions for the salinity, turbidity, and DOC matrix challenges are presented in figure 2. Final test concentrations of turbidity and DOC were slightly below the stated target levels, and there was noticeable settling of turbidity at the highest addition level, but confirm the overall challenge conditions being tested.




Figure 2. *Top Panel:* In situ salinity measured by EXO2 sonde in the laboratory salinity challenge covering ambient plus 3 salinity ranges. *Middle Panel:* In situ turbidity measured by EXO2 sonde (teal) and on grab samples by a Hach 2100 Turbidimeter (olive) during the laboratory turbidity challenge covering ambient plus 2 additions. *Bottom Panel:* In situ fDOM measured by EXO2 sonde (orange) and DOC of discrete samples (dark red) during the DOC challenge covering ambient plus 2 additions.

Results of the laboratory temperature challenge at 5 °C are shown in figure 3. The absolute difference between instrument and reference measurement across all timepoints for trials C2 – C4 ranged from -0.0026 to 0.0170 mgP/L, with a mean of 0.010 ±0.005 mgP/L. The means for each trial are given in Table 3. Measurement differences were significantly higher at 5 °C versus 20 °C for each of three concentration test. The increased offset was 0.007, 0.005 and 0.017 mgP/L, respectively, for C2, C3, and C4. Similar to test results at 20 °C, there was a much larger offset at the C4 level compared to the lower two concentrations for the 5 °C test.

Table 3. Summary of accuracy results for temperature trials assessed by absolute difference (mgP/L) and percent error between instrument and reference measurements.

| Trial | Reference | Systea-PO4 | Absolute Diff | % Error |
|-------|-----------|------------|---------------|-------------|
| C2 | 0.0109 | 0.0199 | 0.0090 | 82.6 |
| C3 | 0.0547 | 0.0601 | 0.0054 | 9.9 |
| C4 | 0.1040 | 0.1178 | 0.0138 | 13.3 |

Results of the laboratory salinity challenge, performed at the C3 concentration level, are shown in figure 4. The absolute difference between instrument and reference measurement across all timepoints for the three added salinity levels ranged from 0.0030 to 0.0333 mgP/L, with a mean of 0.0205 ±0.0078 mgP/L. The means for each salinity trial are given in Table 4. The zero salinity results are taken from the initial concentration challenge on day 1. There was a statistically significant response to increased salinity with the offsets increasing (over-predicted) as salinity increased. A linear regression of the measurement differences versus salinity was significant ($p < 0.001$; $r^2 = 0.86$) with a slope of 0.0009 and intercept of -0.0010. The average offset at salinity 30 was around 0.028 mgP/L higher than for the zero salinity trial which corresponded to a relative error of approximately 90%.

Table 4. Summary of accuracy results for salinity trial assessed by absolute difference (mgP/L) and percent error between instrument and reference measurements.

| Trial | Reference | Systea-PO4 | Absolute Diff | % Error |
|-------|-----------|------------|----------------|-------------|
| 0 | 0.0621 | 0.0619 | -0.0002 | 0.3 |
| 10 | 0.0443 | 0.0580 | 0.0137 | 31.0 |
| 20 | 0.0385 | 0.0584 | 0.0199 | 51.6 |
| 30 | 0.0297 | 0.0575 | 0.0278 | 93.4 |

Results of the laboratory turbidity challenge, performed at the C3 concentration level, are shown in figure 5. The absolute difference between instrument and reference measurement across all timepoints for the two added turbidity levels ranged from -0.0208 to 0.0046 mgP/L, with a mean of -0.0079 ± 0.0091 mgP/L. The means for each turbidity trial are given in Table 5. Results for the zero turbidity level are taken from the initial concentration challenge on day 1. There is no known reason for the large difference in accuracy for the NTU100 trial, but the measurement offset was 0.016 mgP/L lower than that observed at 0 and 10 NTU trials. A linear regression of the measurement differences versus turbidity was significant ($p < 0.01$; $r^2 = 0.59$), with a slope of -0.0008 and intercept of 0.004, however the trend line was clearly forced by the large decrease at 100 NTU and test results should not be interpreted to suggest a strong predictable relationship.

Table 5. Summary of accuracy results for turbidity trials assessed by absolute difference (mgP/L) and percent error between instrument and reference measurements.

| Trial | Reference | Systea-PO4 | Absolute Diff | % Error |
|-------|-----------|------------|----------------|-------------|
| 0 | 0.0621 | 0.0619 | -0.0002 | 0.3 |
| 10 | 0.0525 | 0.0524 | -0.0001 | 0.1 |
| 100 | 0.0520 | 0.0363 | -0.0157 | 30.1 |

Results of the laboratory DOC challenge, performed at the C3 concentration level, are shown in figure 6. The absolute difference between instrument and reference measurement across all timepoints for the two added DOC levels ranged from -0.0018 to 0.0100 mgP/L, with a mean of 0.0018 ± 0.0034 mgP/L. The means for each of the DOC trials are given in Table 6. Results for the zero DOC level are taken from the initial concentration challenge on day 1. A linear regression of the measurement differences versus DOC concentration was not significant ($p = 0.71$; $r^2 = 0.01$).

Table 6. Summary of accuracy results for Laboratory testing assessed by absolute difference (mgP/L) and percent error between instrument and reference measurements for each individual trial condition within each matrix challenge.

| Trial | Reference | Systea-PO4 | Absolute Diff | % Error |
|-------|-----------|------------|----------------|------------|
| 0 | 0.0621 | 0.0619 | -0.0002 | 0.3 |
| 1 | 0.0560 | 0.0594 | 0.0034 | 6.1 |
| 10 | 0.0762 | 0.0763 | 0.0002 | 0.2 |

Lab Temperature Challenge




Figure 3. *Top Panel:* Plot of instrument (blue dots) and reference (red dots) measurements of PO₄ (mgP/L) in the temperature response challenge covering concentration ranges C2 – C4 measured at 5 °C test conditions. Five replicate measurements were made at each concentration level along with one measurement at ambient level. *Bottom Panel:* Plot of the absolute difference between Systea-PO4 and reference measurement.

Lab Salinity Challenge




Figure 4. *Top Panel:* Plot of instrument (blue dots) and reference (red dots) measurements of PO_4 (mgP/L) at four salinity levels for the C3 concentration. Five replicate measurements were made at each concentration level. *Bottom Panel:* Plot of the absolute difference between Systea- PO_4 and reference measurement.

Lab Turbidity Challenge




Figure 5. *Top Panel:* Plot of instrument (blue dots) and reference (red dots) measurements of PO₄ (mgP/L) at three turbidity levels for the C3 concentration. Five replicate measurements were made at each concentration level. *Bottom Panel:* Plot of the absolute difference between SysTea-PO4 and reference measurement.

Lab DOC Challenge




Figure 6. *Top Panel:* Plot of instrument (blue dots) and reference (red dots) measurements of PO_4 (mgP/L) at three DOC levels for the C3 concentration. Five replicate measurements were made at each concentration level. *Bottom Panel:* Plot of the absolute difference between Systea- PO_4 and reference measurement.

A summary of measurement differences between the Systea-PO4 and reference sample for each trial of each laboratory challenge is presented in figure 7. It is not known why the measurement offset greatly increased for the C5 concentration range trial, which corresponded to a test level of 0.41 mgP/L. There was a significant temperature effect in measurement accuracy with the Systea-PO4 over-predicting concentrations at 5 °C more than at 20 °C. There was also a significant salinity effect, with concentrations being over-predicted with increasing salinity. The effects of added turbidity or DOC were less clear, with no real predictable pattern. Results of measurement differences averaged across all trials within each of the challenge matrices are presented in Table 7.




Figure 7. Global summary of difference between instrument and reference measurements for all laboratory tests at each trial conditions for the Systea-PO4 analyzer.

Table 7. Measurement differences in mgP/L (min, max, mean, stdev) between instrument and reference concentrations averaged across all trials within a laboratory challenge.

| Systea-PO4 | Range | Temp | Salinity | Turbidity | DOC |
|-------------------|---------------|---------------|---------------|----------------|---------------|
| min | -0.0043 | 0.0054 | 0.0137 | -0.0157 | -0.0002 |
| max | 0.0215 | 0.0138 | 0.0278 | -0.0001 | 0.0034 |
| mean | 0.0046 | 0.0094 | 0.0205 | -0.0079 | 0.0018 |
| stdev | 0.0099 | 0.0042 | 0.0071 | 0.0110 | 0.0023 |

RESULTS of FIELD TESTS

Moored field tests were conducted to examine the performance of the Systea-PO4 to consistently track natural changes in PO₄ over extended field deployments with durations of 31-84 days. In addition, field tests examined the reliability of the instrument, i.e., the ability to maintain integrity or stability of data collection over time. Reliability was determined by quantifying the percent of expected data that was recovered and useable. The performance of the Systea-PO4 was examined in three separate field tests at various ACT Partner sites to include a range of biogeochemical conditions. The range and mean for temperature and salinity for each test site is presented in Table 8. The reference temperature and conductivity data was measured by RBR thermistors and a SeaBird SBE 26 or Xylem EXO2 sonde that were mounted at the same sampling depth as the test instrument. The Systea-PO4 was calibrated and programmed for deployment by the manufacturer representative.

Table 8. Range and average for temperature, and salinity at each of the test sites during the sensor field deployments. Temperature and salinity were measured by RBR temperature loggers and a SeaBird SBE 26 or a Xylem EXO2 mounted on the instrument rack or in the tank for the duration of the deployment.

| SITE (deployment period/duration) | | Temperature (° C) | Salinity (PSU) |
|---|------|-------------------------------|--------------------------|
| Maumee River | Min. | 20.1 | 0.0 |
| 26May – 27Jun | Max. | 27.7 | 0.3 |
| (n = 32 days) | Mean | 24.3 | 0.2 |
| | | | |
| Chesapeake Bay | Min. | 20.0 | 12.7 |
| 18Jul – 10Oct | Max. | 31.1 | 16.9 |
| (n = 84 days) | Mean | 27.2 | 14.7 |
| | | | |
| Kaneohe Bay | Min. | 24.5 | 27.3 |
| 3Oct – 2Nov | Max. | 27.9 | 34.8 |
| (n = 31 days) | Mean | 26.3 | 34.2 |
| | | | |

Deployment at Maumee River Bowling Green, Ohio

A 32 day deployment occurred from May 26 through June 27 in the Maumee River, at the facilities of the Bowling Green, Ohio Water Treatment Plant (Figure 8). The deployment site was located at 41.48° N, 83.74° W, in a flow-through tank located in the water treatment plant pump house. The pump house is located above the Maumee, approximately 200 m up river from the water treatment intake and approximately 35 km from the Maumee outflow into Lake Erie. River water was continuously pumped into a 180 gallon test tank where it was mixed using two submerged pumps. The residence time in the tank was approximately 10 minutes. The instrumentation was suspended within the tank with the sampling inlet 0.2 m off the bottom.



Figure 8. Aerial view of the Maumee River (left) and the flow through deployment tank (right).

Time series results of ambient conditions for river discharge, temperature, specific conductivity, turbidity and chlorophyll are given in figure 9. Temperature ranged from 20.5 – 27.7°C, specific conductivity from 423 - 689 $\mu\text{S}/\text{cm}$, turbidity from 8 – 681 NTU, and chlorophyll from 4.5 – 131 $\mu\text{g}/\text{L}$ over the duration of the field test.

The Systea-PO4 operated successfully during the entire 32 day deployment sampling at 15 minute intervals. The Systea-PO4 generated 3031 observations out of a possible 3052 for a data completion result of 99.3%. During the deployment 21 data points were not reported by the instrument. Time series results of the Systea-PO4 measurements and corresponding reference PO₄ results are given in figure 10 (top panel). PO₄ measured by the Systea-PO4 ranged from -0.0039 to 0.1773 mgP/L compared to a range of 0.001 to 0.118 mgP/L within the reference samples.




Figure 9. Environmental conditions encountered during the 32 day freshwater deployment in the Maumee River at Waterville, OH. *Top Panel:* Variation in river discharge over the term of the deployment. *Middle Panel:* Variation in temperature (green) and Conductivity (red) at the depth of the sensors, measured by an EXO 2 Sonde. *Bottom Panel:* Time series of turbidity (blue) and chlorophyll (dark yellow) as measured by the EXO 2 Sonde. The large spike in turbidity (681 NTU) was produced during a nutrient addition test when sediment accumulated on the bottom was stirred up from additional mixing of the tank.

The time series of the difference between instrument and reference PO_4 measurements for each matched pair ($n=50$ of a possible 51 observations) is given in the bottom panel of figure 10. One of the 51 possible comparisons was lost because of missing instrument data. The average and standard deviation of the measurement difference over the total deployment was 0.004 ± 0.024 mgP/L with a total range of -0.086 to 0.035 mgP/L. There was no significant trend in measurement difference over time as estimated by linear regression ($p= 0.86$; $r^2=0.001$).




Figure 10. *Top Panel:* Time series plot of the Systea-PO₄ measurement (blue dots) and reference measurements (red dots) of phosphate in mgP/L. *Bottom Panel:* Time series plot of the difference between the Systea-PO₄ and reference measurements of phosphate in mgP/L (instrument – reference) during the freshwater deployment in the Maumee River at Waterville, OH.

A cross-plot of all matched observations for the deployment is given in figure 11. The linear regression of instrument versus reference measurement was significant ($p < 0.01$; $r^2 = 0.25$) but with a slope of only 0.28 and intercept of 0.019. Measurement accuracy clearly declined when concentrations exceeded 0.05 mgP/L.




Figure 11. Maumee River field response plot for the 32 day deployment of the Systea-PO₄ compared to reference PO₄ samples. The plotted line represents a 1:1 correspondence, the blue line represents the linear regression.

Photographs of test instrument before and after the field deployment to indicate potential impact of biofouling (Figure 12).



Figure 12. Photographs of the Systea-PO4 prior to and following a 32 day field test in the Maumee River.

Deployment at Chesapeake Biological Laboratory (CBL)

An 84 day moored field test was conducted in Chesapeake Bay from July 18 to October 10, 2016. The deployment was located at 38.32°N, 76.45°W attached to the side of a floating pier at the mouth of the Patuxent River (Figure 13.) The site was brackish with an average water depth of 2.2 m at the test site.




Figure 13. Aerial view of CBL deployment site (left) and instrument deployment rack off dock during deployment (right).

Time series results of ambient conditions for tidal height, temperature, salinity, turbidity and chlorophyll are given in figure 14. Temperature ranged from 20.0 to 31.3°C, salinity from 12.7 to 16.9 PSU, turbidity from 0.5 to 936.3 NTU and chlorophyll from 0.2 to 97.1 $\mu\text{g/L}$ over the duration of the field test.

The Systea-PO4 operated continuously for the period of its deployment sampling at 30 minute intervals but was retrieved 12 days prior to the scheduled end date of the deployment to send to the next field test in HI. (Note: It was originally intended that a second unit would be used in HI, but the manufacturer was not able to secure an additional unit in time.) While the unit was deployed it reported 3086 of a possible 3402 accepted values for a data completion result of 90.7%. During its operation, 64 values were flagged by the instrument with no data, and 252 were omitted as outliers when reported values exceeded 10 times above observed levels or were less than -0.004. Time series results of the Systea-PO4 and corresponding reference PO₄ results are given in figure 15 (top panel). For the interval deployed, the range of accepted values reported by the Systea-PO4 was -0.004 to 0.068 mgP/L, compared to 0.003 to 0.034 mgP/L within reference samples.

The bottom panel of figure 15 presents the time series of the difference between the Systea-PO4 and reference PO₄ for each matched pair (n=78 comparisons out of a total of 103, (22 missing data points from early retrieval, 2 data points were flagged, and 1 data point was omitted as an outlier). The average and standard deviation of the measurement difference for the deployment was 0.019 ± 0.014 mgP/L, with the total range of differences between -0.019 to 0.049 mgP/L.

There no significant trend in measurement difference over time as estimated by linear regression ($p=0.58$; $r^2=0.004$) over the deployment period.




Figure 14. Environmental conditions encountered during the 84 day CBL floating dock deployment. Test sensor array deployed at 1 m fixed depth, variation in local tidal heights indicate active water flow around instrument (*Top Panel*). Variation in temperature (green) and salinity (red) at depth of instrument sensor detected by an EXO2 sonde and two RBR Solo thermistors (*Middle Panel*). Variation in turbidity (blue) and chlorophyll (dark yellow) at depth of instrument sensor detected by an EXO2 sonde (*Bottom Panel*).




Figure 15. Time series of PO_4 measured by the Syste-a- PO_4 during the 84 day CBL field trial. *Top Panel:* Continuous PO_4 recordings from instrument (blue circles) and PO_4 of adjacent grab samples (red circles). *Bottom Panel:* The difference in measured PO_4 relative to reference samples (Instrument mgP/L – Reference mgP/L) observed during deployment.

A cross-plot of the matched observations for the deployment is given in figure 16. A linear regression of the data was not significant ($p=0.58$; $r^2 = 0.004$) and the instrument generally over-predicted concentrations.




Figure 16. CBL field response plot for Systea-PO4 compared to reference PO₄ samples. The plotted line represents a 1:1 correspondence.

Photographs of the Systea-PO4 before and after the field deployment to indicate potential impact of biofouling (Figure 17).



Figure 17. Photographs of the Systea-PO4 instrument prior to and following the CBL field trial.

Deployment off Coconut Island in Kaneohe Bay, Hawaii

A one month long moored field test was conducted in Kaneohe Bay from October 3, 2016 to November 2, 2016. The deployment site was located at 21.43° N, 157.79° W, on a floating dock anchored off Coconut Island (HIMB) in a depth of approximately 16 meters (Figure 18). Kaneohe Bay, located on the eastern side of Oahu, Hawaii, is a complex estuarine system with a large barrier coral reef, numerous patch reefs, fringing reefs, and several riverine inputs. Tides in Kaneohe Bay are semi-diurnal with mean tidal amplitude of approximately 68 cm day.



Figure 18. Aerial view of HIMB deployment site (left) and instrument rack in-situ (right).

Time series results of ambient conditions for tidal height, temperature, and salinity are given in figure 19. Temperature at the sensor level ranged from 24.5 to 27.9 °C and salinity from 27.3 to 34.8 PSU over the duration of the field test

The Systea-PO₄ was deployed on day four of the field test and operated successfully for the remaining 26 days of the deployment. Time series results of the Systea-PO₄ and corresponding reference PO₄ results are given in figure 20 (top panel). During the deployment, the Systea-PO₄ returned 2477 acceptable instrument measurements of a possible 2496 measurements for a data completion result of 99%. During the deployment 13 data points were reported as missing values and 6 observations were omitted as outliers (values >10x observed maximum). The range of values reported by the Systea-PO₄ analyzer was -0.004 to 0.012 mgP/L, compared to the range within reference samples of 0.0024 to 0.0061 mgP/L. The bottom panel of figure 20 presents the time series of the measurement difference between the Systea-PO₄ and reference PO₄ for each matched pair. The average and standard deviation of the differences between instrument and reference readings (n=63 out of a possible 63) were -0.0004 ± 0.002 mgP/L, with a total range in the differences of -0.0058 to 0.0044 mgP/L. There was a small but statistically significant trend in the measurement difference over time ($p=0.0003$; $r^2 = 0.247$) during the deployment, with a slope of 0.0001 mgP/L/d.




Figure 19. Environmental conditions encountered during the one month HIMB deployment on a floating dock off Coconut Island Test sensor array deployed at 1 m fixed depth, variation in local tidal heights indicate active water flow around instrument (*Top Panel*). Variation in temperature (green) and Salinity (red) at depth of instrument sensor detected by an SBE 26 and two RBR Solo thermistors (*Middle Panel and Bottom Panel*).




Figure 20. *Top panel:* Time series of PO₄ measured by the Systea-PO4 deployed during the one month HIMB field trial. Continuous PO₄ recordings from instrument (blue dots) and PO₄ of adjacent grab samples (red circles.) *Bottom Panel:* Time series of the difference between the Systea-PO4 and reference measurements for each matched pair (Instrument mgP/L – Reference mgP/L).

A cross-plot of the matched observations for the deployment is given in figure 21. A linear regression of the data was highly significant but with a low regression coefficient ($p=0.001$; $r^2 = 0.209$). The regression had a slope of 0.999 and y intercept of -0.0004. Under and over predictions were fairly evenly distributed.




Figure 21. HIMB field response plot of Systea-PO₄ compared to reference PO₄ samples. The plotted line represents a 1:1 correspondence, the blue line represents the linear regression.

Photographs of and example of the test instrument prior to deployment and the test instrument after the HIMB field deployment to indicate potential impact of biofouling (Figure 22).



Figure 22. Photographs of the Systea-PO4 prior to and following the one month HIMB field trial.

A global summary of instrument versus reference readings for all three field deployment sites are plotted in figure 23. The Systea-PO4 response varied substantially for the freshwater test compared to the estuarine and marine tests. Overall, it did not accurately measure concentrations greater than 0.04 mgP/L observed during the Maumee River deployment. A linear regression of all the data was significant but with a low regression coefficient ($p < 0.0001$; $r^2 = 0.2$). The slope of the regression was 0.017 with an intercept of 0.496. The global data comparison covered a field concentration range of 0.001 to 0.118 mgP/L.




Figure 23. Global response plot for the Systea-PO4 observed during the three ACT field trials. Black dotted line represents a 1:1 correspondence, the blue line represents the linear regression.

QUALITY ASSURANCE AND QUALITY CONTROL

All technology evaluations conducted by ACT comply with its Quality Management System (QMS), which includes the policies, objectives, procedures, authority, and accountability needed to ensure quality in work processes, products, and services. A QMS provides the framework for quality assurance (QA) functions, which cover planning, implementation, and review of data collection activities and the use of data in decision making, and quality control. The QMS also ensures that all data collection and processing activities are carried out in a consistent manner, to produce data of known and documented quality that can be used with a high degree of certainty by the intended user to support specific decisions or actions regarding technology performance. ACT's QMS meets U.S. Environmental Protection Agency quality standards for environmental data collection, production, and use, and the requirements of ISO/IEC 17025:2005(E), *General requirements for the competence of testing and calibration laboratories*.

An effective assessment program is an integral part of a quality system. The ACT Quality Assurance (QA) Manager independently conducted Technical Systems Audits (TSA) of field tests at Maumee River field trial during May 25-28, 2016, a TSA of the Laboratory test at the Chesapeake Biological Laboratory during July 10-18, 2016 and a data quality review of the reference data sets from all tests conducted during the Nutrient Challenge.

Technical System Audits

A TSA is a thorough, systematic, on-site qualitative audit of sampling and measurement processes and procedures associated with a specific technology evaluation. The objectives of the TSAs conducted during this evaluation were to assess and document the conformance of on-site testing procedures with the requirements of the Test Protocols, the ACT Quality Assurance Project Plan (QAPP), and associated Standard Operating Procedures (SOPs).

The TSA was conducted in accordance with the procedures described in n EPA's *Guidance on Technical Audits and Related Assessments for Environmental Data Operations (EPA QA/G-7)* and ISO 19011, *Guidelines for Quality and/or Environmental Management Systems Auditing*. A TSA checklist based on the Test Protocols was prepared prior to the audits and reviewed by the ACT Director and Senior Scientist. The TSA assessed ACT personnel, the test and analytical facilities, equipment maintenance and calibration procedures, sample collection, analytical activities, record keeping, and QC procedures. Reference sample handling and chain-of-custody by NASL were observed during the laboratory test at CBL.

During the audits, the QA Manager met with ACT technical staff involved in the evaluation and asked them to describe the procedures followed. All procedures were observed; and logbooks, data forms, and other records were reviewed.

Key components of the audit included:

- Assessment of Quality Assurance/Quality Control:
 - Adequacy of procedures, and
 - Adherence to procedures.
- Assessment of Sample System:
 - Sample collection,
 - Analytical procedures, and
 - Documentation.

- Assessment of Data and Document Control:
 - Chain of custody, and
 - Documentation.

The TSAs' findings were positive. The field and laboratory tests were implemented consistent with the Test Protocols, QAPP, and SOPs. Minor deviations were documented in laboratory records. There were no deviations which may have had an effect on data quality for the test. All phases of the implementation of the tests reviewed during the audits were acceptable and performed in a manner consistent with ACT data quality goals. The overall quality assurance objectives of the test were met.

ACT personnel are well-qualified to implement the evaluation and demonstrated expertise in pertinent procedures. Communication and coordination among all personnel was frequent and effective. Internal record keeping and document control was well organized. The ACT staff understands the need for QC, as shown in the conscientious development and implementation of a variety of QC procedures.

All samples were collected as described in the Test Protocols and SOPs. Examination of maintenance and calibration logs provided evidence of recent and suitable calibration of sampling and analytical equipment.

Data Quality

Data Verification, Validation, and Assessment.

Data review is conducted to ensure that only sound data that are of known and documented quality and meet technology evaluation quality objectives are used in making decisions about technology performance. Data review processes are based in part on two EPA guidance documents: *Guidance on Environmental Data Verification and Data Validation (QA/G-8)* [EPA, 2002] and *Guidance on Technical Audits and Related Assessments for Environmental Data Operations (QA/G-7)* [EPA, 2000].

The data were verified and validated to evaluate whether the data have been generated according to the Test Protocols and satisfied acceptance criteria. Data verification evaluates the completeness, correctness, and consistency of the data sets against the requirements specified in the Test Protocols, measurement quality objectives (MQOs), and any other analytical process requirements contained in SOPs.

The ACT QA Manager reviewed the reference data sets from all field and laboratory tests. The number of reference samples collected at each site and the laboratory tests are in Table 10. A total of 346 reference samples were collected for the field and laboratory tests. The overall reference data set included 3,666 distinct analyses.

Table 9. The number of reference samples collected during the laboratory test and at each field site.

| Site | No. of Samples | No. of Replicates per Sample | No. of Analytes ^{1/} Measured in Each Replicate | No. of Measurements |
|---|----------------|------------------------------|--|---------------------|
| Maumee River | 61 | 3 | 3 | 549 |
| CBL – Field | 120 | 3 | 3 | 1080 |
| CBL – Lab | 92 | 5 | 3 | 1380 |
| Hawaii | 73 | 3 | 3 | 657 |
| Total | 346 | | | 3,666 |
| ^{1/} NO ₂ ; NO ₃ ; PO ₄ | | | | |

The data review verified that the sampling and analysis protocols specified in the Test Protocols were followed, and that the ACT measurement and analytical systems performed in accordance with approved methods, based on:

- The raw data records were complete, understandable, well-labeled, and traceable;
- All data identified in the Test Protocols were collected;
- QC criteria were achieved; and
- Data calculations were accurate.

Data validation uses the outputs from data verification and included inspection of the verified field and laboratory data to determine the analytical quality of the data set. A representative set of approximately 10% of the reference data was traced in detail from 1) raw data from field and laboratory logs, 2) data transcription, 3) data reduction and calculations, to 4) final reported data. Validation of the data sets established:

- Required sampling methods were used;
- Sampling procedures and field measurements met performance criteria; and
- Required analytical methods were used.

The data validation also confirmed that the data were accumulated, transferred, summarized, and reported correctly. There is sufficient documentation of all procedures used in the data collection and analysis to validate that the data were collected in accordance with the evaluation’s quality objectives.

A Data Quality Assessment (DQA) is the third and final process of the overall data assessment. It is a scientific and statistical evaluation of validated data to determine if the data are of the right type, quality, and quantity to support conclusions on the performance of the technologies. The DQA determined that the test’s data quality objectives, described in Section 7.1 of the Test Protocols and Section 3.4 and Appendix B of the ACT QAPP (ACT, 2016), were achieved. This evidence supports conclusions that:

- The sampling design performed very well and is very robust with respect to changing conditions.

- Sufficient samples were taken to enable the reviewer to see an effect if it were present.

Audit of Data Quality.

The ACT QA Manager conducted an Audit of Data Quality (ADQ) on verified data to document the capability of ACT’s data management system to collect, analyze, interpret, and report data as specified in the Test Protocols, QAPP, and SOPs. The ADQ determined that the data were accumulated, transferred, reduced, calculated, summarized, and reported correctly. There is sufficient documentation of all procedures used in the data collection and analysis to verify that the data have been collected in accordance with ACT quality objectives.

Table 10. Results of Field Duplicates (FD) for the Maumee River mooring test.

| Date/Time | Rep | PO ₄ | Mean | Std Dev | ABS Diff | CV% |
|---------------|-----|-----------------|-------|---------|----------|-------|
| 6-16-16 9:00 | FD1 | 0.0094 | 0.008 | 0.0016 | 0.002 | 19.2 |
| | FD2 | 0.0072 | | | | |
| 6-17-16 12:00 | FD1 | 0.0071 | 0.008 | 0.0005 | 0.001 | 7.11 |
| | FD2 | 0.0079 | | | | |
| 6-20-16 10:00 | FD1 | 0.0122 | 0.012 | 0.0003 | 0.0005 | 2.59 |
| | FD2 | 0.0117 | | | | |
| 6-23-16 11:00 | FD1 | 0.0203 | 0.020 | 0.0001 | 0.0001 | 0.489 |
| | FD2 | 0.0202 | | | | |

Table 11. Results of Field Duplicates (FD) for the Chesapeake Bay, MD mooring test.

| Date/Time | Rep | PO ₄ | Mean | Std Dev | ABS Diff | CV% |
|---------------|-----|-----------------|--------|---------|----------|-------|
| 7-20-16 10:00 | FD1 | 0.0040 | 0.0039 | 0.0002 | 0.0003 | 6.1 |
| | FD2 | 0.0037 | | | | |
| 7-26-16 14:00 | FD1 | 0.0042 | 0.0039 | 0.0005 | 0.0007 | 12.09 |
| | FD2 | 0.0036 | | | | |
| 8-2-16 10:00 | FD1 | 0.0057 | 0.0057 | 0.0000 | 0.0001 | 0.827 |
| | FD2 | 0.0057 | | | | |
| 8-10-16 16:00 | FD1 | 0.0131 | 0.0148 | 0.0024 | 0.0034 | 16.14 |
| | FD2 | 0.0164 | | | | |

| | | | | | | |
|----------------|-----|--------|--------|--------|--------|-------|
| 8-23-16 12:00 | FD1 | 0.0199 | 0.0197 | 0.0003 | 0.0004 | 1.56 |
| | FD2 | 0.0194 | | | | |
| 9-8-16 10:00 | FD1 | 0.0224 | 0.0249 | 0.0035 | 0.0050 | 14.11 |
| | FD2 | 0.0274 | | | | |
| 9-16-16 12:00 | FD1 | 0.0189 | 0.0195 | 0.0008 | 0.0011 | 4.00 |
| | FD2 | 0.0200 | | | | |
| 10-4-16 14:00 | FD1 | 0.0157 | 0.0144 | 0.0019 | 0.0027 | 13.27 |
| | FD2 | 0.0130 | | | | |
| 10-10-16 10:00 | FD1 | 0.0216 | 0.0221 | 0.0007 | 0.0010 | 3.10 |
| | FD2 | 0.0225 | | | | |

Table 12. Results of Field Duplicates (FD) for the Kaneohe Bay, HI mooring test

| Date/Time | Rep | NO₃ | Mean | Std Dev | ABS Diff | CV % |
|------------------|------------|-----------------------|-------------|----------------|-----------------|-------------|
| 10-6-16 14:00 | FD1 | 0.0035 | 0.0036 | 0.000 | 0.0001 | .664 |
| | FD2 | 0.0036 | | | | |
| 10-12-16 11:00 | FD1 | 0.0034 | 0.0033 | 0.0002 | 0.0003 | 5.77 |
| | FD2 | 0.0031 | | | | |
| 10-17-16 9:00 | FD1 | 0.0035 | 0.0034 | 0.0001 | 0.0001 | 2.07 |
| | FD2 | 0.0034 | | | | |
| 10-26-16 9:00 | FD1 | 0.0039 | 0.0040 | 0.0002 | 0.0003 | 4.68 |
| | FD2 | 0.0042 | | | | |
| 11-1-16 9:00 | FD1 | 0.0053 | 0.0053 | 0.0001 | 0.0002 | 2.25 |
| | FD2 | 0.0052 | | | | |

Table 13. Results of Field Trip Blanks all deployments.

| Maumee River | | Chesapeake Bay | | Kaneohe Bay | |
|-------------------------|------------------------------|-----------------------|------------------------------|--------------------|------------------------------|
| Field Blank ID | PO ₄ (Std Dev) | Field Blank ID | PO ₄ (Std Dev) | Field Blank ID | PO ₄ (Std Dev) |
| GLFB1 | 0.0008 (0.0001) | CBLFB1 | 0.0027 (0.0001) | HIFB1 | 0.0017 (0.0000) |
| GLFB2 | 0.0012 (0.0003) | CBLFB2 | 0.0026 (0.0001) | HIFB2 | 0.0016 (0.0002) |
| GLFB3 | 0.0021 (0.0001) | CBLFB3 | 0.0014 (0.0001) | HIFB3 | 0.0013 (0.0002) |
| GLFB4 | 0.0027 (0.0004) | CBLFB4 | 0.0011 (0.0003) | HIFB4 | 0.0013 (0.0002) |
| -- | -- | -- | -- | HIFB5 | 0.0010 (0.0001) |
| Mean (Std Dev) | 0.0017 (0.001) | Mean (Std Dev) | 0.002 (0.001) | Mean (Std Dev) | 0.001 (0.0003) |
| Grand Mean (Std Dev) | | | | | 0.002 (0.0007) |

ACKNOWLEDGEMENTS:

We wish to acknowledge the support of all those who helped plan and conduct the verification test, analyze the data, and prepare this report. In particular we would like to thank our Technical Advisory Committee, Dr. Suzanne Bricker, National Oceanic and Atmospheric Administration, Dr. Brian Pellerin, U.S. Geological Survey, Dr. Dwane Young, U.S. Environmental Protection Agency, Dr. Matt Cohen, University of Florida, Dr. R. David Holbrook, National Institute for Standards and Technology, Mr. Chris Gross, U.S. Department of Agriculture NRCS, Dr. Joe Rudek, Environmental Defense Fund for their advice and direct participation in various aspects of this evaluation. Earle Buckley also provided critical input on all aspects of this work and served as the independent Quality Assurance Manager. This work has been coordinated with, and funded by, the National Oceanic and Atmospheric Administration, Integrated Ocean Observing System program.

June 1, 2017


Date



Approved By: Dr. Mario Tamburri
ACT Executive Director

June 1, 2017


Date



Approved By: Dr. Tom Johengen
ACT Chief Scientist

June 1, 2017

Date



Approved By: Dr. Earle Buckley
Quality Assurance Supervisor

System Comments on the Nutrient Challenge WIZ performances and results

We would like to bring to your attention the importance of the proper choice of the sample filtering system to increase the quality and the stability of the in situ measurements. During the last 25 years of continuous development of Nutrient sensors, we have also done wide investigations on the online/in situ sample filtration solutions. We have developed, in fact, several specific filtering solutions, depending on the matrix of the sample to be measured, and the requested parameter. We may choose the most adequate filtering solution based on sample matrix and in particular:

1-Waste Water;

For this matrix we normally apply a Fast Loop sampling, a 25Microns SS tubular filter, on board air compressor to generate air blasts inside the filter and remove the particles attached to the filter during the filtration.

2-Surface Water/Sea Water low turbidity;

For this matrix we use a 25microns SS tubular filter, with 100 square centimeters of filtering surface, and small internal dead volume(10ml). The filter surface is antifouling protected with a copper coil. The sampling line, from the filter to the analyzer, is back washed with acidic solution. This type of filtration has been used during the 3 WIZ deployments of the Nutrient Challenge.

3-River Water high turbidity & thin sediments;

For this matrix we normally suggest a 0.1 Microns hollow fibers filter, complete with Copper sampling probe and on board back wash. The filter can be used, coupled with our WIZ-Log, in river water with high sediments content for several weeks unattended. The filtrate water is microfiltered and then totally transparent and usable also for sensors cleaning. In case of any doubt about the possible sediments content in Surface or Sea Water, this type of filter can be the safest solution, but one must consider that there is a need of dedicated DL, to handle the filtration sequence.

-TP/TN/COD/TOC in Waste Water;

We use the Fast Loop sampling plus 100Microns SS tubular filter complete with on board air compressor to generate air blasts to clean the filter. For the above parameters in not allowed a filtration below 100Microns, because we need to provide not only the dissolved fractions but also the fractions in the Organic Matter bound to the sediments.

After the 3 WIZ deployments experiences, we should give the following filtering suggestions:

Maumee River: The filtering solution 3 should be the right solution.

CBL: Filter 2 will be ok; Filter 3 better if DL available.

Hawaii: Filter 2 OK; Filter 3 better if DL available.